

## **QUICK START GUIDE**

ActaStim-S Spine Fusion Stimulator






#### **Assembly & Application**



B Electrodes 2 will be placed on the lower back on either side of the site of spine fusion surgery area. Clean and dry the skin before attaching the electrodes.



Connect Generator Cable (1a) or (1b) to Generator (3).



Insert Battery 4 into Generator 3.
This turns on device including display/audio/vibrate. Check back page for indicators.



Generator 3 can be clipped on belt, carried in a pocket or bag, or placed on furniture.



### **Important Safety Information**



**Do not use Generator in the shower.** (Disconnect Electrodes from Generator Cable)



See patient manual for additional safety information.

#### **Charger Operation**

Connect Charging Cable 5 to Power Supply Block 6.

Connect Charging Cable 5 to Battery Charger 7.









# Get started with ActaStim Sync

- Scan QR code or search "ActaStim Sync" in the AppStore or Google Play
- Download ActaStim Sync app
- Open app and follow the prompts



#### **Indicators & Troubleshooting**

#### **Indicator** Action Description Press once, twice or Button to Mute three times to turn off Audio/Vibrate audio and vibrate, or to and Turn Off Display toggle display on/off. **Green Battery** = Charged No action required. Green Check = Treating 1. Check connections Orange Electrodes = A and B. Connection Issue 2. Check that Battery may be green, electrodes (i) are in orange or flashing contact with the skin. 3. Check manual. Orange Battery = Need Check that Spare to make sure that Spare Battery is charging Battery is charging as or start charging current battery is low Spare Battery. **Green Check** = Treating Orange Flashing Remove battery **Battery** = Need to from Generator change Battery and replace with No Green Check = charged Battery Not Treating from Charger. Orange Triangle = Product Fault Battery may be green, Call Customer Support orange or flashing 1-800-901-5667. No Green Check = Not Treating Orange Flashing 270 days of treatment complete. Battery, Triangle Device will no longer provide treatment. and Electrodes = Contact Physician. End of Product Life